

Michelangelo

Suggested Coding Options¹

January 2020

The Michelangelo Hand is used on a custom fabricated prosthesis. Due to customization, additional coding may apply.

Base Code Options (includes Socket, Forearm, Humeral Section, Shoulder Bulkhead, Cables, Batteries, Charger, and Switch/Myoelectric Control)

Transcarpal switch/myo	L6026
Wrist Disarticulation switch/myo	L6920/L6925
Below Elbow switch/myo	L6930/L6935
Elbow Disarticulation switch/myo	L6940/L6945
Above Elbow switch/myo	L6950/L6955
Shoulder Disarticulation switch/myo	L6960/L6965
Interscapular-Thoracic switch/myo	L6970/L6975

Socket Options (add as appropriate, not all inclusive)

Test Sockets	L6680 (WD/BE), L6682 (ED/AE), L6684 (SD/IT)
Suction Socket	L6686
Frame Socket	L6687 (WD/BE), L6688 (ED/AE), L6689 (SD), L6690 (IT)
Ultralight Material	L7400 (WD/BE), L7401 (ED/AE), L7402 (SD/IT)
Acrylic	L7403 (WD/BE), L7404 (ED/AE), L7405 (SD/IT)

Replacement Socket

Replacement Socket	L6883 (BE/WD), L6884 (ED/AE), L6885 (SD/IT)
--------------------	---

Elbow Options

AxonArm Ergo 12K501=M /M1 /M2	L6693 (forearm counterbalance) + L6638 (elec.lock) + L7499* (MP Control of the APS System) <i>*See: Michelangelo Billing Tips</i>
---	--

AxonArm Hybrid 12K500=M /M1 /M2	L6693 (forearm counterbalance)
---	--------------------------------

Harness (if used)

Above Elbow Harness, Triple Control, Steel 21A35=1	L6677 (harness triple) + L6655 (standard ctrl cable, extra)
--	---

¹ The product/device “Supplier” (defined as an O&P practitioner, O&P patient care facility, or DME supplier) assumes full responsibility for accurate billing of Ottobock products. It is the Supplier’s responsibility to determine medical necessity; ensure coverage criteria is met; and submit appropriate HCPCS codes, modifiers, and charges for services/products delivered. It is also recommended that Supplier’s contact insurance payer(s) for coding and coverage guidance prior to submitting claims. Ottobock Coding Suggestions and Reimbursement Guides are based on reasonable judgment and are not recommended to replace the Supplier’s judgment. These recommendations may be subject to revision based on additional information or alpha-numeric system changes.

Michelangelo

Suggested Coding Options¹

January 2020

Additional Switches (if used)

9X14 Harness Pull Switch	L6611 (additional switch)
9X18 Cable Pull Switch	
9X25 Rocker Switch	
9X37 Pressure Switch	
9X50 & 9X52 Linear Transducer	

Hand, Hook Options – Above Elbow Fitting

Michelangelo Hand 8E500	L7499* (Michelangelo Hand) + L6882 (MP control) + L6881 (Autograsp) + L6629 (QD) + L6890 (glove) + Wrist Option (below) <i>*See Michelangelo Bill Tips</i>
-----------------------------------	--

Michelangelo Wrist Options

9S501 Passive Rotation	L6624 (flex/ext rot wrist)
OR	OR
9S503 Active Rotation	L6624 (flex/ext wrist) + L7259 (rotator) + L6882 (MP Control)

Michelangelo Transcarpal Hand 8E550=*	L7499* (Michelangelo hand) + L6882 (MP control) + L6881 (Autograsp) + L6890 (glove) <i>*See Michelangelo Billing Tips</i>
---	---

AxonHook 8E600=*	L7499* (AxonHook) + L6882 (MP control) + L6881 (Autograsp) + L6629 (QD) + Michelangelo Wrist Option (above) <i>*See Michelangelo Bill Tips</i>
----------------------------	--

Liner options (most common, not all inclusive)

14Y5 Silicone Arm Liner with CVD coating	L6694 (locking liner)
14A1 Lock Set	L6698 (lock)
6Y416 Custom Shape Plus UE Urethane Liner (model made by hand)	For Initial Liner use: L6696 (custom liner congenital/atypical traumatic) or L6697 (custom liner not congenital/atypical traumatic)
6Y417 Custom Shape Plus Urethane UE Replacement Liner (same mold)	For Replacement Liner use: L6694 (locking liner) or L6695 (liner not for lock)

¹ The product/device “Supplier” (defined as an O&P practitioner, O&P patient care facility, or DME supplier) assumes full responsibility for accurate billing of Ottobock products. It is the Supplier’s responsibility to determine medical necessity; ensure coverage criteria is met; and submit appropriate HCPCS codes, modifiers, and charges for services/products delivered. It is also recommended that Supplier’s contact insurance payer(s) for coding and coverage guidance prior to submitting claims. Ottobock Coding Suggestions and Reimbursement Guides are based on reasonable judgment and are not recommended to replace the Supplier’s judgment. These recommendations may be subject to revision based on additional information or alpha-numeric system changes.

Michelangelo

Suggested Coding Options¹

January 2020

6Y455 Custom Liner UE	For Initial use: L6696 (custom liner congenital/atypical traumatic) or L6697 (custom liner not congenital/atypical traumatic) For other than initial use: L6694 (locking liner) or L6695 (liner not for lock)
#617S1=H/W(2,3,4,5,6,8,10) Bocklite for a socket insert (padding)	L6691 (Removable insert, each)
88L1=* Custom Silicone TR Liner	Contact the Ottobock Reimbursement Team for help 800.328.4058 / reimbursement911@ottobock.com
Replacement Glove (prefabricated)	
8S500 /01 / 02 / 11/ AxonSkin Gloves	L6890 (glove any material)
Replacement Electrodes (initial electrodes & cables included in base code)	
13E202=60 Suction Socket Electrode (60hz)	Included in base code, use L7510 for replacement
13E129 Electrode Cable	Included in base code, use L7510 for replacement
Replacement Battery/Charger (initial battery and charger are included in the base code)	
757B500 / 757B501 AxonEnergy Integral	L7367 LITHIUM ION BATTERY, REPLACEMENT (check with payer, may require modifiers)
757L500 AxonCharge Integral	L7368 LITHIUM ION BATTERY CHARGER, REPLACEMENT (check with payer, may require modifiers)

Ottobock North America, Reimbursement
 P 800 328 4058
professionals.ottobockus.com/Reimbursement
reimbursement911@ottobock.com

¹ The product/device “Supplier” (defined as an O&P practitioner, O&P patient care facility, or DME supplier) assumes full responsibility for accurate billing of Ottobock products. It is the Supplier’s responsibility to determine medical necessity; ensure coverage criteria is met; and submit appropriate HCPCS codes, modifiers, and charges for services/products delivered. It is also recommended that Supplier’s contact insurance payer(s) for coding and coverage guidance prior to submitting claims. Ottobock Coding Suggestions and Reimbursement Guides are based on reasonable judgment and are not recommended to replace the Supplier’s judgment. These recommendations may be subject to revision based on additional information or alpha-numeric system changes.