

# Instructions for Use

**50K326=1** Post-Op Knee Brace, Cool

## Application Instructions



**Step 1:** Lay the brace open on a flat surface. Extend the hinges so they line up on the side of the patient's leg.




**Step 2:** After measuring the patient's leg, remove the pull tabs and locate the Velcro clasps at the end of each strap.



**Step 3:** Remove the Velcro clasp and trim the straps to the desired length using scissors.



**Step 4:** Determine the length of the hinges, and locate and unlock the 4 gray dials on the sides of the brace to extend the length.



**Step 5:** Set the desired flexion and extension stops to the degrees needed for the patient. Once the correct settings are achieved, lock the hinges on both sides into position.



**Step 6:** Completed fitting! Have your patient remove and apply the brace themselves using the quick-release buckles.

## Care Instructions

Hand wash with mild detergent and air dry.

## Warranty

The 50K326 is covered by a 6 month warranty against all mechanical or manufacturer defects, from the date of purchase.

### Distributed by Ottobock HealthCare LP

11501 Alterra Parkway Suite 600 . Austin, TX 78758  
 Ottobock US . P 800 328 4058 . F 800 962 2549 . professionals.ottobockus.com  
 Ottobock Canada . P 800 665 3327 . F 800 463 3659 . professionals.ottobock.ca

**Warning:** Read all instructions and warnings prior to use. Do not remove brace unless instructed by your medical treatment professional. Do not change device settings without consulting a medical professional. This device will not prevent or reduce all injuries. Proper rehabilitation and activity modification are also an essential part of a safe treatment program. Consult with your medical treatment professional regarding safe and appropriate activity level while wearing this device. If you experience pain, swelling, skin irritation, or any other reactions while using this product, immediately consult your healthcare provider. The hinge on this brace is designed to limit and/or control range of motion. It is not designed to stabilize your knee when you are weight-bearing or take the place of a walking aid. Follow your healthcare provider's advice regarding weight-bearing activities and always use a proper assistance device, such as crutches or a walker.

**Caution:** This device is RX only. Federal law restricts this device to sale by or on the order of a licensed health care practitioner. This device is for single patient use only.

## Sizing

- Measure leg circumference 6" above mid-patella.
- Fits thigh circumferences up to 38".

## Universal

- 50K326=1 Post-Op Knee Brace, Cool

## Indications for Use

For locked or limited range of motion control following injury or operative procedures of the knee.

# Instructions for Use

**50K326=2** Post-Op Knee Brace, Full Foam

## Application Instructions




**Step 1:** Lay the brace open on a flat surface. Extend the hinges so they line up on the side of the patient's leg.



**Step 2:** Wrap the foam pads around the patient's thigh and calf, and mark any extra foam. Trim the foam to the desired size for the patient's leg.



**Step 3:** Line the telescoping hinges of the brace so they are centered on each side of the leg, and adjust to the correct height. Once complete, Velcro the hinges to the foam pads, remove the pull tabs, and attach the straps on both sides.



**Step 4:** Determine the length of the hinges, and locate and unlock the 4 gray dials on the sides of the brace to extend the length.



**Step 5:** Set the desired flexion and extension stops to the degrees needed for the patient. Once the correct settings are achieved, lock the hinges on both sides into position.



**Step 6:** Completed fitting! Have your patient remove and apply the brace themselves using the quick-release buckles.

### Care Instructions

Hand wash with mild detergent and air dry.

### Warranty

The 50K326 is covered by a 6 month warranty against all mechanical or manufacturer defects, from the date of purchase.

### Distributed by Ottobock HealthCare LP

11501 Alterra Parkway Suite 600 . Austin, TX 78758  
 Ottobock US . P 800 328 4058 . F 800 962 2549 . professionals.ottobockus.com  
 Ottobock Canada . P 800 665 3327 . F 800 463 3659 . professionals.ottobock.ca

**Warning:** Read all instructions and warnings prior to use. Do not remove brace unless instructed by your medical treatment professional. Do not change device settings without consulting a medical professional. This device will not prevent or reduce all injuries. Proper rehabilitation and activity modification are also an essential part of a safe treatment program. Consult with your medical treatment professional regarding safe and appropriate activity level while wearing this device. If you experience pain, swelling, skin irritation, or any other reactions while using this product, immediately consult your healthcare provider. The hinge on this brace is designed to limit and/or control range of motion. It is not designed to stabilize your knee when you are weight-bearing or take the place of a walking aid. Follow your healthcare provider's advice regarding weight-bearing activities and always use a proper assistance device, such as crutches or a walker.

**Caution:** This device is RX only. Federal law restricts this device to sale by or on the order of a licensed health care practitioner. This device is for single patient use only.

### Sizing

- Measure leg circumference 6" above mid-patella.
- Fits thigh circumferences up to 38".

### Universal

- 50K326=2 Post-Op Knee Brace, Full Foam

### Indications for Use

For locked or limited range of motion control following injury or operative procedures of the knee.